

Interactive  
Comment

## ***Interactive comment on “Spectrophotometric high-precision seawater pH determination for use in underway measuring systems” by S. Aßmann et al.***

**S. Aßmann et al.**

steffen.assmann@hzg.de

Received and published: 29 July 2011

For a better comparison of the underway system to the bench-top system we improved the temperature control. Now a stable temperature is reached before the CRM runs out. We carried out repetitive measurements ( $N = 14$ ) and found an average offset of 0.0005 pH units compared to the bench-top spectrometer.

Underway system:  $8.0974 \pm 0.0006$

Bench-top system:  $8.0969 \pm 0.0004$

Interactive comment on Ocean Sci. Discuss., 8, 1339, 2011.

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

Discussion Paper

